



INDRAPRASTHA GAS LIMITED  
IGL Bhawan, 4, Community Center,  
Sector 9, R.K.Puram, New Delhi-110022  
Email: sachin.mujoo@igl.co.in/ sweeti.kumari@igl.co.in  
Phone: +91 11 4607 4682

TENDER DOCUMENT NO.IGL/ET2/CP/CC18291  
Procurement of Master mass flow meters (Prover system) for calibration of Dispensers and LCV flow meters

REPLY TO BIDDERS QUERIES

Sr No.	Section	Page No.	Tender Clause	Bidder Query	IGL Remarks
1	Section V, Scope of work	61	<b>Prover Inlet Process Connections</b> 1. NZS receptacle (OPW make) for Car. 3. Quick connect coupling (OPW make) for LCV Hose.	(1) Bidder is requesting to kindly approve OEM make NZS receptacle for Car. We have already supplied many provers with OEM make receptacle. PTR will be shared during detailed engineering. (3) Bidder is requesting to kindly approve Parker/Eaton/OEM make QRC coupling and also confirm QRC size.	For point 1, <b>Tender condition prevails</b> For point 3, Bidder may supply OPW/Parker make quick connect coupling for LCV hose. And Size of quick connect coupling stem end is 1/2"
2	Section V, Scope of work	61	<b>Prover Outlet Process Connections:</b> 1 Quick connect coupling stem end (OPW make) suitable for Car, Bus and LCV Hose 2, 3 & 4 Outlet Hose (Swagelok/Parker make)	(1) Bidder is requesting to kindly approve Parker/Eaton/OEM make QRC coupling and please confirm the QRC Size and Hose Size require for LCV. (2, 3 & 4) Bidder is requesting to kindly approve Zec SPA/Transfer Oil/Eaton make Hoses. These hoses make are also approved by many CGDs Companies. Kindly consider the same.	For point 1, Bidder may supply Parker/OPW make quick connect coupling for LCV hose. Size of quick connect coupling stem end is 3/4" suitable for car, bus and LCV hose. Size of LCV hose : Hose ID- 1/2" For point 2,3 & 4: Bidder may supply Swagelok/Parker/Eaton make outlet hose.
3	Section V, Scope of work	61	SS Fitting - Swagelok/Parker make	Bidder is requesting to kindly approve Dk LOK/Astec fitting make SS fittings. DK-LOK/Astec make are also approved by many CGDs Companies. Kindly consider the same.	* <b>Tender condition prevails</b>
4	Section V, Scope of work	61	SS Tubing - Sandvik/Tubacex/Parker make	Bidder is requesting to kindly approve Ratnamani/Jindal Saw make SS Tubes. Ratnamani/Jindal Saw make are also approved by many CGDs Companies. Kindly consider the same.	Bidder may use Sandvik/Tubacex/Parker/Ratnamani SS tubing

Note:

1. Please note that only the clarifications of the queries received from the bidder's are replied in this corrigendum, under reply to bidder's Query. In case, if any discrepancies is observed between the tender document and the clarification stated above, the tender condition shall prevails for those cases.
2. All other terms & conditions of tender document remain unaltered.
3. Bidders are requested to visit IGL website / IGL's tendering website <https://igl.ewizard.in> regularly for further announcement
4. Please submit the signed and stamped copy of this corrigendum along with your offer.