



INDRAPRASTHA GAS LIMITED
IGL Bhawan, 4, Community Center,
Sector 9, R.K.Puram, New Delhi-110022
Email: sachin.mujoo@igl.co.in/ sweeti.kumari@igl.co.in
Phone: +91 11 4607 4682

TENDER DOCUMENT NO.IGL/ET2/CP/CC17990
RATE CONTRACT FOR SUPPLY OF PRESSURE, TEMPERATURE GAUGES & TRANSMITTERS FOR METERING SKID, IN-HOUSE COMPRESSORS, DISPENSERS
AND CASCADES INSTALLED IN CNG STATIONS.

REPLY TO BIDDERS QUERIES

Sr No.	Section	Tender Clause	Page No.	Description	Bidder Query	IGL Remarks
1	VI	SOR	65-66	Schedule of Rates	In Temperature Gauge specification: Dial size 100 mm – ½” NPT (M) connection & 63 mm – 3/8” BSP (M) connection mentioned but in SOR TG items 5,6,8 & 9 only 63 mm dial mentioned in description. So, kindly confirm which dial size required item wise.	Description as per SOR is correct, all TG (line item no. 5,6,8 & 9) dial Size will be 63mm. For specification please refer to Corrigendum-III .
2	VI	SOR	65-66	Schedule of Rates	In SOR item no. 10, Differential Gauge, 100 mm, 1 bar detailed specification not mentioned in scope of work.	For detailed specification please refer to Corrigendum-III .
3	VI	SOR	65-66	Schedule of Rates	In SOR item no. 15, Pressure Transmitter required range not mentioned.	For Line item 15: Model mentioned in SOR (WA-10) For technical specification refer SOW Point 1 (Pressure transmitter)
4	V	SOW	60-63	Scope of Work	The query is in Technical Specifications of Capillary Temperature gauge IGL has mentioned "Response time : With Thermo-well < 50 sec, without Thermo-well < 20 sec".	As per SOW for Temperature gauges only response time was mentioned for the equipment i.e. (Response time : With Thermo-well < 50 sec, without Thermo-well < 20 sec). Thermowell is not required for the said equipment but equipment quoted should match the response time mentioned in SOW.

Note:

1. Please note that only the clarifications of the queries received from the bidder's are replied in this corrigendum, under reply to bidder's Query. In case, if any discrepancies is observed between the tender document and the clarification stated above, the tender condition shall prevails for those cases.
2. All other terms & conditions of tender document remain unaltered.
3. Bidders are requested to visit IGL website / IGL's tendering website <https://petroleum.ewizard.in> regularly for further announcement
4. Please submit the signed and stamped copy of this corrigendum along with your offer.