



INDRAPRASTHA GAS LIMITED
IGL Bhawan, 4, Community Center,
Sector 9, R.K.Puram, New Delhi-110022
Email: prateek.shrivastava@igl.co.in/ gaurav.agrawal1@igl.co.in/ dnagar@igl.co.in
Phone: +91 11 4607 4651

TENDER DOCUMENT NO.IGL/ET2/CP/CM18145
PROCUREMENT OF PORTABLE GAS DETECTOR

REPLY TO PREBID QUERIES

S.NO	Section	Tender Clause.	Page No.	Description	Bidder Query	IGL Reply
1	V	Scope of Work Clause no. 2 & 3	60	2. The detector should have sensor for LEL & Oxygen Measurement : LEL (combustible gas): IR Type Oxygen: Electrochemical/Electrolyte Sensor should be over range protected. The detector should be manufactured recently not more than six months old (at the time of supply). 3. Range: LEL- 0-100 % LEL (calibrated to methane) resolution/ Accuracy 1%, Oxygen: minimum range 0-25%, resolution accuracy 0.1%.	With reference to the subject tender, where your requirement is detector having Oxygen and IR type Combustible where the gas range for IR type combustible (CH4) required in 0-100% LEL. If you need the Combustible IR sensor to be calibrated in CH4, then gas measuring range will be in 0-100% Vol. 0-100% LEL reading is possible with the sensor calibrated with Propane. Kindly consider this mail as our request to review your technical specification and make the amendment accordingly which will allow us to participate in this tender.	Tender Conditions Prevail